



Math-Net.Ru

All Russian mathematical portal

T. Cheng, Ch. Tao, Super-resolution microscopy based on interpolation and wide spectrum de-noising,
Computer Optics, 2023, Volume 47, Issue 4, 614–619

<https://www.mathnet.ru/eng/co1162>

Use of the all-Russian mathematical portal Math-Net.Ru implies that you have read and agreed to these terms of use

<https://www.mathnet.ru/eng/agreement>

Download details:

IP: 18.97.14.86

May 21, 2025, 08:45:12



Super-resolution microscopy based on interpolation and wide spectrum de-noising

T. Cheng¹, T. Chenchen¹

¹ Guangxi University of Science and Technology,
545006, P.R. China, Liuzhou, Chengzhong District, Avenue Donghuan, 268

Abstract

In the conventional single-molecule localizations and super-resolution microscopy, the pixel size of a raw image is approximately equal to the standard deviation of the point spread function. Such a raw image is referred to herein as a conventional raw image, based on which better single molecule localization effect and efficiency can be achieved. It is found that both interpolation and de-noising can effectively improve the Signal to Noise Ratio of the conventional raw image. The conventional raw image, the de-noised, the interpolated and the de-noised interpolated are compared and analyzed and compressed sensing is used for super-resolution reconstruction. The simulation results show that both the highest Signal to Noise Ratio and the best super-resolution reconstruction can be obtained by de-noising the interpolated conventional raw image. This method also renders the best super-resolution reconstruction and minimum gradient in the real experiment. De-noising the interpolated conventional raw image is an effective method to improve the super-resolution microscopy.

Keywords: super-resolution microscopy; interpolation; de-noising; point spread function; compressed sensing.

Citation: Cheng T, Chenchen T. Super-resolution microscopy based on interpolation and wide spectrum de-noising. *Computer Optics* 2023; 47(4): 614-619. DOI: 10.18287/2412-6179-CO-1272.

Introduction

Optical microscopy imaging has several advantages such as magnification, real-time visualization and non-invasive observation, making optical microscope a basic tool for cell research. Owing to the diffraction phenomenon, the point light source forms a diffuse spot (i.e., an Airy disc) on the focal plane, which is the point spread function (PSF) of the microscope. The diffraction limit restricts an optical microscope to resolve the subcellular organization and cell organelles. However, super-resolution microscopy (SRM) imaging based on fluorescent probes can observe the inside of cells and subcellular life activities. It breaks the diffraction limit and increases the imaging resolution [1–6].

Super-resolution microscopy (SRM) techniques include single-molecule localization microscopy (SMLM), stimulated emission depletion microscopy (STED), structured illumination microscopy (SIM) and super-resolution optical fluctuation imaging (SOFI), etc., among which SMLM achieves the highest resolution. SMLM methods include (fluorescence) photoactivated localization microscopy ((F)PALM), stochastic optical reconstruction microscopy (STORM), and DNA-based point accumulation for imaging in nanoscale topography (DNA-PAINT), etc[1–5, 7–9]. Since thousands frames of raw images in the same field of view are acquired for single molecule localization in SMLM, the temporal resolution of SMLM is very poor. Fluorescent molecules (fluorophores or emitters) must be sparsely activated in SMLM. If PSFs of some fluorescent molecules overlap, these fluorescent molecules will not be efficiently localized [3, 4, 10–14].

Compressed sensing (CS) can reconstruct the raw image with high density of fluorescent molecules. PSFs of fluorescent molecules can overlap in CS, greatly reducing the number of frames of the raw image required to reconstruct the final super-resolution image. Thus the temporal resolution is much better than SMLM [13–16]. The CS mathematical model between a raw image and a super-resolution image is shown in Eq. (1). The vectors, \mathbf{y} and \mathbf{x} consist of column-wise concatenations of the raw image and the super-resolution image (i.e., pixelated original image) respectively. The measurement matrix \mathbf{A} is subject to the PSF of the microscope. The acquired raw image corresponds to the i^{th} column of \mathbf{A} if only one molecule emits fluoroscopic photons at the position index i of \mathbf{x} [13–16].

$$\min\|\mathbf{x}\|_0 \quad \text{s.t.} \quad \mathbf{y} = \mathbf{A}\mathbf{x}, \quad (1)$$

where $\mathbf{x} \in \mathbb{R}^N$, $\mathbf{y} \in \mathbb{R}^M$, $\mathbf{A} \in \mathbb{R}^{M \times N}$, $M < N$, M and N are natural numbers. \mathbf{x} is a vector containing N elements. \mathbf{y} is a vector containing M elements. \mathbf{A} is a matrix of size $M \times N$. $\min\|\mathbf{x}\|_0$ is a objective function. $\mathbf{y} = \mathbf{A}\mathbf{x}$ is a constraint function. If the objective function is $\min\|\mathbf{x}\|_1$ rather than $\min\|\mathbf{x}\|_0$, Eq. (1) is transformed to a convex optimization problem.

In conventional SMLMs, if the pixel size of the raw image is approximately equal to the standard deviation (s.d.) of the PSF, good and fast localization can be achieved. The existing research results show that interpolation can also improve the CS reconstruction [13, 14, 17]. Interpolation based on the bicubic method can make the pixel size of the raw image become half of the original's [14].

Noise in the raw image is inevitable. Especially in low light environment, noise can almost drown the signal.

The noises of a raw image mainly include a shot noise following a Poisson distribution, a readout noise following a Gaussian distribution and a background. The reconstruction of CS is also restricted by noise [4, 12, 17–19]. Large noise can even lead to reconstruction failure. Wide spectrum denoising (WSD) is a SRM denoising algorithm for various random noises, which helps improve the reconstruction of CS [18, 20]. Hereinafter, the raw image whose pixel size is approximately equal to the s.d is referred to as a conventional raw image (CR). CI is the interpolated CR. CW and CIW respectively represent the de-noised CR and CI with WSD.

The denoising effect and CS reconstruction of CIW are studied in this paper, and they also compared with those of CR, CW and CI. It is found that both interpolation and de-noising can effectively improve the Signal to Noise Ratio (SNR) of the CR and the highest SNR can be obtained by de-noising the CI. The simulation and real experimental results show that the best super-resolution reconstruction can be obtained by de-noising the CI. Thus de-noising the CI is an effective method to improve the super-resolution microscopy.

1. Wide spectrum de-noising and quantitative evaluation

WSD can effectively remove random noise such as Poisson and Gaussian noise from very low density to ultra-high density fluorescent molecular distribution scenarios. If the measurement matrix \mathbf{A} is operated by orthogonalization and normalization, the measurement matrix \mathbf{A}_0 can be obtained. Row orthogonal normalization is an operation in the matrix theory. It makes the rows of the matrix completely orthogonal. Moreover, it makes the 2-norm of each row of the matrix equal to 1. Through \mathbf{A}_0 and \mathbf{A} , the operator matrix \mathbf{T} which is equivalent to the row orthogonal normalization operation can be obtained, where $\mathbf{T} = \mathbf{A}_0 \mathbf{A}^T (\mathbf{A} \mathbf{A}^T)^{-1}$. Therefore, $\mathbf{y} = \mathbf{A} \mathbf{x}$ can be equivalently converted to $\mathbf{T} \mathbf{y} = \mathbf{T} \mathbf{A} \mathbf{x}$. If a singular value decomposition is applied to \mathbf{T} , $\mathbf{T} = \mathbf{U} \mathbf{S} \mathbf{V}^T$. \mathbf{S} is a diagonal sparse matrix that is composed of singular values of \mathbf{T} . \mathbf{V}^T is a transpose matrix of \mathbf{V} . Therefore, $\mathbf{T} \mathbf{y} = \mathbf{T} \mathbf{A} \mathbf{x}$ can be equivalently converted to $\mathbf{S} \mathbf{V}^T \mathbf{y} = \mathbf{S} \mathbf{V}^T \mathbf{A} \mathbf{x}$. Let $\mathbf{z} = \mathbf{S} \mathbf{V}^T \mathbf{y}$. \mathbf{z} 's elements greater than cri are set to cri . The new \mathbf{z} is represented by \mathbf{z}_{WSD} . The threshold value cri is the largest absolute values of elements of \mathbf{z} between the indices $M \times 0.9$ and $M \times 0.95$. The de-noised raw image is \mathbf{y}_{WSD} , $\mathbf{y}_{WSD} = \mathbf{T}^{-1} \mathbf{U} \mathbf{y}_{WSD}$ [18, 20].

SNR can indicate the overall quality of the de-noised raw image and the reconstructed super-resolution image [14]. Structural similarity index measure (SSIM) can measure the similarity between two images. The SSIM values range between 0 and 1. If the two images are identical, the value of SSIM is equal to 1 [10, 21].

$$SNR = 20 \times \log_{10} \left(\frac{\|\mathbf{x}\|_2}{\|\mathbf{x}_R - \mathbf{x}\|_2} \right), \quad (2)$$

where \mathbf{x} is a true signal, $\mathbf{x} \in \mathbf{R}^N$; $\|\cdot\|_2$ is the norm of a vector; and \mathbf{x}_R denotes the reconstructed signal corresponding \mathbf{x} .

$$SSIM(x, y) = \frac{(2\mu_x \mu_y + c_1)(2\sigma_{xy} + c_2)}{(\mu_x^2 + \mu_y^2 + c_1)(\sigma_x^2 + \sigma_y^2 + c_2)}, \quad (3)$$

where μ_x and σ_x represent the mean and the variance of image x , respectively. σ_{xy} represents the covariance of image x and y . $c_1 = (k_1 L)^2$ and $c_2 = (k_2 L)^2$. L is the dynamic range of pixel values. $k_1 = 0.01$, $k_2 = 0.03$.

2. Interpolation and de-noising of simulated conventional raw images

In real experiments, an inverted optical microscope with a $100\times$ oil immersion objective lens (N.A. 1.40) was used. Tubulin of HeLa cells was stained with Alexa-647. The fluorescence wavelength was 670 nm. The fluorescent lights were acquired using an EMCCD. Twenty CR frames were collected. The PSF was a Gaussian function. The full width at half maximum (FWHM) of the PSF was 239.286 nm. The same parameters were used in simulation experiments.

To be distinguished from the pixels of the CR, the pixels of the super-resolution image were referred to as grids. The super-resolution image's grid was $1/8$ of the pixel size of the CR. If the size of the CR was 7×7 pixels, it was located in the middle of the 64×64 grid super-resolution image. The pixel size of the CI was $1/2$ of the pixel size of the CR.

To evaluate the de-noising performance of WSD for different molecular densities and pixel sizes for CR and CI, simulated CRs with known real molecular positions were generated. The simulation randomly placed K molecules in a 64×64 grid region. K represented the sparsity in CS. K ranged from 1 to 16, with corresponding molecular densities from 1.869 to $30 \mu\text{m}^2$. The grid size was 11.43 nm. The effective pixel size of the CR, 45.714 nm, matched the pixel size of the microscope setup and was roughly equal to half of the s.d. of the PSF of the setup. The size of the CR was 7×7 pixels in the middle of the 64×64 grid super-resolution image. The simulation was for a photon number of 3,000 per molecule and a uniform background of 16 photons per pixel. Poisson noise and Gaussian noise (Gaussian noise variances of 0.01 and 0.0025) were added to each CR frame. The size of the CI was 13×13 pixels.

Fig. 1 is mean SNRs of 500 frames of CR and CI before and after de-noising. For each K , the simulation was repeated 500 times as shown in Fig. 1. CI is the CI. CW and CIW respectively represent the CW and CI with WSD.

The de-noising performance analysis curves are shown in Fig. 1a and b. Both interpolation and de-noising can effectively improve the SNR of the CR. The highest SNR can be obtained by de-noising the CI. In Fig. 1a for the high noise environment (Gaussian noise with a variance of 0.01), the SNR of the CI is lower than that of the CW. In Fig. 1b for the low noise environment (Gaussian

noise with a variance of 0.0025), the SNR of the CI is almost the same as that of the CW.




Fig. 1. Comparison of mean SNRs based on 500 simulated CRs before and after interpolation and de-noising. The simulation is for a photon number of 3,000 per molecule and a background of 64 photons per pixel. Poisson noise and Gaussian noise are added to each frame of the CR. The y axis is labeled as SNR (dB). The x axis is labeled as both molecular density and signal sparsity. a) Gaussian noise variance of 0.01, b) Gaussian noise variance of 0.0025




Fig. 3. Comparison of the CR (Gaussian noise with a variance of 0.01), the CW, the CI, and the CIW. The numbers indicate SNR (dB) of the images a) CR 14.421 dB, b) CW 16.906 dB, c) CI 16.366 dB, d) CIW 18.18dB




Fig. 4. Comparison of the simulated noise-free CRs, the corresponding simulated noise-free CI, and the true super-resolution image, Scale bars: 1 μm . a) noise-free CR, b) noise-free CI, c) True Super

Figs. 2 and 3 are one simulation of Fig. 1a at K equal to 4. Fig. 2a is the true super-resolution image containing four fluorescence molecules. Fig. 2b is the noise-free CR corresponding to Fig. 2a. Fig. 2c is the noise-free CI.




Fig. 2. Comparison of the true super-resolution image, the corresponding simulated noise-free CR and the corresponding simulated noise-free CI. a) 4 molecules, b) noise-free CR, c) noise-free CI

Fig. 3a is the CR after adding background, Poisson noise and Gaussian noise with variance of 0.01 to Fig. 2b. Fig. 3b is the CW. Fig. 3c and d are respectively interpolated Fig 3a and de-noised Fig. 3c. SNRs of the CR and CI after de-noising is improved by 2.485 dB and 1.814 dB respectively.

3. Simulation data analysis

To evaluate the CS reconstruction for CIW, 20 frames of simulated super-resolution image with known true molecular positions and their corresponding CR frames were generated. The simulation randomly placed molecules in a grid region. The molecular density was $6.304 \mu\text{m}^{-2}$.

Fig. 4a is a CR frame without noise. Fig. 4b is a CI frame without noise. Fig. 4c is the true super-resolution image.




Fig. 5. Comparison of the simulated (interpolated) CR (Gaussian noise with a variance of 0.01) and the corresponding super-resolution images before and after de-noising. The upper row is one frame of CR and CI before and after de-noising. The lower row is the final super-resolution images corresponding to the upper row. The number pairs indicate SSIM and SNR (dB) of the images. Scale bars: 1 μm . a) CR (0.312, 7.059 dB), b) CW (0.452, 11.397 dB), c) CI (0.25, 8.615 dB), d) CIW (0.325, 10.791 dB), e) CR, f) CW (0.574, -2.093 dB), g) CI, h) CIW (0.671, -2.021 dB)




Fig. 6. Comparison of the simulated (interpolated) CR (Gaussian noise with a variance of 0.0025) and the corresponding super-resolution images before and after de-noising. The upper row is one frame of CR and CI before and after de-noising. The lower row is the final super-resolution images corresponding to the upper row. The number pairs indicate SSIM and SNR (dB) of the images. Scale bars: 1 μm . a) CR (0.45, 12.261 dB), b) CW (0.547, 15.86 dB), c) CI (0.373, 13.873 dB), d) CIW (0.433, 15.977 dB), e) CR, f) CW (0.637, -1.76 dB), g) CI (0.779, -1.68 dB), h) CIW (0.796, -1.411 dB)

Fig. 5a is one CR frame after adding background, Poisson noise and Gaussian noise with variance of 0.01 to Fig. 4a. Fig. 5b is the CW frame corresponding to Fig. 5a. Fig. 5c and d respectively are the CI and the CIW frames. SNRs of Fig. 5b and d increased by 4.338 dB and 2.176 dB than Fig. 5a and c. SSIMs of Fig. 5b and d increased by 0.14 and 0.075 than Fig. 5a and c.

The lower row of Fig. 5 is the final super-resolution image reconstructed by CS. SSIM and SNR of Fig. 5h reached the maximum, respectively 0.671 and -2.021 dB. The cell microtubule structure is clearer and thinner. SSIM and SNR increased by 0.097 and 0.072 dB respectively compared with Fig. 5f. The CS reconstruction of Fig. 5e and g failed. No valid microtubule structures can be seen. Although some scattered structures can be seen in Fig. 5g, the result is too poor to be useful. The microtubule structure of the rectangle area in Fig. 5h is cleaner than that

in Fig. 5f. All three microtubules are clearly identifiable in Fig. 5h. However, only two microtubules can be seen clearly in Fig. 5f. Therefore, the CS reconstruction of Fig. 5h is the best.

Fig. 6a is one CR frame after adding background, Poisson noise and Gaussian noise with variance of 0.0025 to Fig. 4a. SNRs of Figs. 6b and d respectively increased by 3.599 dB and 2.104 dB than Figs. 6a and c. SSIMs of Figs. 6b and d respectively increased by 0.097 and 0.06 than Figs. 6a and c.

The second row of Fig. 6 is the final super-resolution image reconstructed by CS. SSIM and SNR of Fig. 6(h) reached the maximum, respectively 0.796 and -1.411 dB. The cell microtubule structure is clearer and thinner. SSIM and SNR increased by 0.159 and 0.349 dB respectively compared with Fig. 6f. SSIM and SNR are better than Fig. 6g, too. The CS reconstruction of Fig. 6e failed.

No valid microtubule structures can be seen. The area between microtubules in the rectangular area in Fig. 6h has almost no discrete points, which is cleaner than that in Fig. 6f and g. Therefore, the CS reconstruction of Fig. 6h is the best.

4. Real experimental data analysis

Fig. 7 shows the real experimental CR and CI before and after de-noising and the corresponding CS reconstruction results. Fig. 7a–d of the upper row is CR and CI frames before and after de-noising. Since the real raw image cannot be known, the sum of the absolute values of the horizontal and vertical gradients of the raw

image is used as an indicator of measuring the de-noising effect. The gradients of CR, CW, CI, and CIW were 7.458×10^5 , 7.146×10^5 , 4.25×10^5 and 3.974×10^5 respectively. Through comparison of the images, it is obvious that the raw image after denoising is smoother than that before denoising, indicating prominent denoising effects.

Comparing the CS reconstruction results of CR, CW, CI and CIW in the lower row, we can see noticeably improved spatial resolutions of CI and CIW and clearer, thinner cell microtubule structure. The super-resolution reconstruction of Fig. 7f is the worst, and Fig. 7g is the second-worst.




Fig. 7. Comparison of the experimental (interpolated) CR and the corresponding final super-resolution images before and after denoising. (a)–(d) of the upper row is one frame of the experimental CR and CI before and after de-noising. The number indicate gradient of the images. (f)–(i) of the lower row is corresponding final super-resolution images before and after denoising, respectively. (e) and (i) is plots of photon-count profiles obtained by measurements made along the red lines in (h)–(i). The maximum wave peak is 4300. Scale bars: 1 μm . a) CR 7.458×10^5 , b) CW 7.146×10^5 , c) CI 4.25×10^5 , d) CIW 3.974×10^5 , e) CI-profile, f) CR, g) CW, h) CI, i) CIW, j) CIW-profile

The peaks in Fig. 7e and j correspond to the photon-count profiles at the red line in Fig. 7h and i. The highest peak value is 4300. The peak in Fig. 7j is thinner and less than Fig. 7e. Therefore, the super-resolution reconstruction of Fig. 7i is the best.

Conclusion

In the conventional single-molecule localizations and super-resolution microscopy, the pixel size of a raw image is approximately equal to the standard deviation of the point spread function. Both interpolation and denoising can effectively improve the SNR of the CR. It is found that the highest SNR and minimum gradient can be obtained by de-noising the CI. In the high noise environment, the SNR of the CI is lower than that of the CW and the CI cannot achieve effective super-resolution reconstruction, while in the low noise environment, the SNR of the CI is almost the same as that of the CW. The simulation and real experimental results show that the best super-resolution reconstruction can be obtained by de-noising the CI. In conclusion, de-noising the CI is an effective method to improve the super-resolution microscopy.

Acknowledgements

The work was funded by Guangxi National Natural Science Foundation (2022GXNSFAA035593), National Natural Science Foundation of China (81660296, 41461082).

References

- [1] Gabitto MI, Marie-Nellie H, Pakman A, Pataki A, Darzacq X, Jordan MI. A Bayesian nonparametric approach to super-resolution single-molecule localization. *Ann Appl Stat* 2021; 15(4): 1742-1766. DOI: 10.1214/21-AOAS1441.
- [2] Nevskiy O, Tsukanov R, Gregor I, Karedla N, Enderlein J. Fluorescence polarization filtering for accurate single molecule localization. *APL Photon* 2020; 5(6): 061302. DOI: 10.1063/5.0009904.
- [3] Costello I, Cox S. Analysing errors in single-molecule localisation microscopy. *Int J Biochem Cell Biol* 2021; 27(2): 105931. DOI: 10.1016/j.biocel.2021.105931.
- [4] Rimoli CV, Valades-Cruz CA, Curcio V, Mavrikis M, Brasselet S. 4polar-STORM polarized super-resolution imaging of actin filament organization in cells. *Nat Commun* 2022; 13(1): 301. DOI: 10.1038/s41467-022-27966-w.

- [5] Kwon J, Elgawish MS, Shim SH. Bleaching-resistant super-resolution fluorescence microscopy. *Adv Sci* 2022; 9(9): 2101917. DOI: 10.1002/advs.202101817.
- [6] Henriques R, Lelek M, Fomasiero EF, Valtorta F, Zimmer C, Mhlanga MM. QuickPALM: 3D real-time photoactivation nanoscopy image processing in ImageJ. *Nat Methods* 2010; 7(5): 339-340. DOI: 10.1038/nmeth0510-339.
- [7] Kozma E, Kele P. Fluorogenic probes for super-resolution microscopy. *Org Biomol Chem* 2019; 15(17): 215-233. DOI: 10.1039/c8ob02711k.
- [8] Chung J, Jeong U, Jeong D, Go S, Kim D. Development of a new approach for low-laser-power super-resolution fluorescence imaging. *Anal Chem* 2021; 101(94): 618-627. DOI: 10.1021/acs.analchem.1c01047.
- [9] Jeong D, Kim D. Super-resolution fluorescence microscopy-based single-molecule spectroscopy. *Bulletin of the Korean Chemical Society* 2022; 43(12): 316-327. DOI: 10.1002/bkcs.12471.
- [10] Roa C, Le VND, Mahendroo M, Saytashev I, Ramella-Roman JC. Auto-detection of cervical collagen and elastin in Mueller matrix polarimetry microscopic images using. *Biomed Opt Express* 2021; 12(4): 2236-2249. DOI: 10.1364/BOE.420079.
- [11] Holden SJ, Uphoff S, Kapanidis AN. DAOSTORM: an algorithm for high-density super-resolution microscopy. *Nat Methods* 2011; 8(4): 279-280. DOI: 10.1038/nmeth0411-279.
- [12] Junhong M, Cédric V, Hagai K, Lina C, Nicolas O, Seamus H, Suliana M, Chul YJ, Michael U. FALCON: fast and unbiased reconstruction of high-density super-resolution microscopy data. *Sci Rep* 2014; 4(4): 4577. DOI: 10.1038/srep04577.
- [13] Zhu L, Zhang W, Elnatan D, Huang B. Faster STORM using compressed sensing. *Nat Methods* 2012; 9(7): 721-723. DOI: 10.1038/nmeth.1978.
- [14] Cheng T, Chen DN, Yu B, Niu HB. Reconstruction of super-resolution STORM images using compressed sensing based on low-resolution raw images and interpolation. *Biomed Opt Express* 2017; 8(5): 2445-2457. DOI: 10.1364/BOE.8.002445.
- [15] Arjoune Y, Kaabouch N, Ghazi HE, Tamtaoui A. A performance comparison of measurement matrices in compressive sensing. *Int J Commun Syst* 2018; 31(2): e3576. DOI: 10.1002/dac.3576.
- [16] Calisesi G, Ghezzi A, Ancora D, D'Andrea C, Valentini G, Farina A, Bassi A. Compressed sensing in fluorescence microscopy. *Prog Biophys Mol Biol* 2022; 60(6): 66-80. DOI: 10.1016/j.pbiomolbio.2021.06.004.
- [17] Thompson RE, Larson DR, Webb WW. Precise nanometer localization analysis for individual fluorescent probes. *Biophys J* 2002; 43(82): 2775-2783. DOI: 10.1016/s0006-3495(02)75618-x.
- [18] Cheng T, Chen DN, Li H. Wide spectrum denoising (WSD) for superresolution microscopy imaging using compressed sensing and a high-resolution camera. *J Phys Conf Ser* 2020; 1651: 012177. DOI: 10.1088/1742-6596/1651/1/012177.
- [19] Beier HT, Ibey BL. Experimental comparison of the high-speed imaging performance of an EM-CCD and sCMOS camera in a dynamic live-cell imaging test case. *PLoS ONE* 2014; 18(9): e84614. DOI: 10.1371/journal.pone.0084614.
- [20] Cheng T. Wide spectrum denoising method for microscopic images. US Patent 16845110 of July 2, 2022.
- [21] Lee G, Oh JW, Her NG, Jeong WK. DeepHCS ++ : Bright-field to fluorescence microscopy image conversion using multi-task learning with adversarial losses for label-free high-content screening. *Med Image Anal* 2021; 70: 101995. DOI: 10.1016/j.media.2021.101995.

Authors' information

Tao Cheng, Doctor of Technical Sciences, Professor, Professor of School of Mechanical and Automotive Engineering of Guangxi University of Science and Technology. Research interests: medical imaging and image processing. He is also the **corresponding author** of this paper. E-mail: ctnp@163.com.

Chenchen Tao, Master's student of School of Foreign Studies of Guangxi University of Science and Technology. Research interests: microscopy and image processing. She is also the corresponding author of this paper. E-mail: taochenchen621@126.com.

Received January 6, 2023. The final version – February 20, 2023.